



OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC
HEALTH OFFICER, KANDHAMAL
E-Mail: cdmophulbani@gmail.com,
Phone / Fax No- 06842-253249

No. 3120 Phulbani Dated 8 / 3 /2019
DWH / 52 / MISC. EIF / 18-19

To

The Director,
I & PR Department, Govt. of Odisha
Lok Sampark Bhawan, Bhubaneswar
E-mail id: ipr.advt@gmail.com / iprenews@gmail.com

Sub: Publication of the advertisement related to supply & installation of Equipment, Instrument, Furniture and Other Items to Kandhamal district.

Sir,

Please find here with a specimen copy of the advertisement related to **supply & installation of Equipment, Instrument, Furniture and Other Items to Kandhamal district** for Publication of the same in two nos. of daily news paper (One time) by Dt.08.03.2019.

This is for favour of your kind information and necessary action.


Yours faithfully,


Chief District Medical & PH Officer,
Kandhamal

Memo No. 3130 / DWH

Date:

1. Copy to the DI & PRO, Kandhamal for information and necessary action.
2. Copy to the DIO, NIC, Kandhamal for information with a request to publish the same along with the enclosures (enclosed herewith) in the District website of Kandhamal district for wide publicity.
3. Copy to the Head Clerk, O/o CDM&PHO, Kandhamal for information.


Chief District Medical & PH Officer,
Kandhamal

OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, KANDHAMAL

No. 3129 /DWH

Tender Call Notice

Date 8-3-19

Sealed tenders are invited from authorized firms for **supply & installation of Equipment, Instrument, Furniture and Other Items to the ZSS, Kandhamal**. The details are available in the district website: www.kandhamal.nic.in . The eligible bidders may submit their tender papers **on or before 18.03.2019 by 5 PM** through Registered Post / Speed Post / Courier only to the undersigned. The **Technical Bid document will open on 19.03.2019 at 11 AM**. Who will qualify in the technical bid, their **Price Bid will be opened on 20.03.2019 at 11 AM** by the purchase committee in the office chamber of the undersigned. The undersigned reserves the right to accept or reject any or all the tender without assigning any reason thereof.

Sd/-

Chief District Medical & PH Officer, Kandhamal



**CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER
KANDHAMAL**

**Tel: 06842-253385;
e-mail :cdmophulbani@gmail.com,**

Tender Reference No. DWH/2018-19/3129

**TENDER DOCUMENT
FOR
SUPPLY & INSTALLATION
OF
MEDICAL EQUIPMENTS, INSTRUMENT,
FURNITURE AND OTHER ITEMS**

**Address for Correspondence- Office of the Chief District
Medical & Public Health Officer, Kandhamal
At/Po-Phulbani, Dist- Kandhamal, Odisha
Pin-762001**

ABSTRACT OF CONTENTS IN THE TENDER PAPER

| SI No | Content | Page No |
|--------------|---|----------------|
| 1 | Section-I Notice Inviting Tender | 01 |
| 2 | Section-II Important Instructions to be Noted Carefully by the Tenderers | 02 |
| 3 | Section-III Terms & Conditions for Supply & Installation of Medical Equipments, Instruments, Furniture and Other Items | 03-05 |
| 4 | Annexure-I Check List | 06 |
| 5 | Annexure-II List of Items to be tendered | 07-12 |
| 6 | Annexure-III List of items quoted Format | 13 |
| 7 | Annexure-IV Price Bid Format | 13 |
| 8 | Annexure-V Declaration by Notary regarding Black Listed | 14 |
| 9 | Annexure-VI Authorization Letter from Manufacturer Unit | 15 |
| 10 | Annexure-VII Self Declaration regarding rate quoted | 16 |
| 11 | Annexure-VIII Self Declaration regarding binding of rate quoted | 16 |
| 12 | Annexure-IX Self Declaration regarding supply of items within stipulated time period | 17 |
| 13 | Annexure-X Declaration regarding minimum turn over | 18 |
| 14 | Annexure-XI Technical Specification of the items | 19-27 |

OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, KANDHAMAL

SECTION -I

NOTICE INVITING TENDER

Tender Reference No. : . DWH/2018-19/3129

Dated: 08.03.2019

TENDERS ARE INVITED FROM ELIGIBLE BIDDERS AS PER THE ELIGIBILITY CRITERIA FOR SUPPLY & INSTALLATION OF MEDICAL EQUIPMENT, INSTRUMENT, FURNITURES AND OTHER ITEMS

| | | |
|---|---|--|
| 1 | Period of Availability of Tender Document | From 09.03.2019 TO 18.03.2019 (Downloadable from website: www.kandhamal.nic.in) In case of any bid amendment and clarification, responsibility lies with the bidders to collect the same from the above mentioned website before last date of submission of tender document and the tender inviting authority shall have no responsibility for any delay / omission on part of the bidder. |
| 2 | Last date & time for submission of Tender | Date: 18.03.2019, Time: 5 PM Address of Submission of Bid: OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER, KANDHAMAL At/Po- Phulbani, Dist.- Kandhamal, Pin- 762001 (Through Speed post / Registered post / Courier) |
| 3 | Date, time and place of opening of Tender | A. Technical Bid (Cover-A): Opening – 19.03.2019 at 11.00 AM in the address mentioned above. B. Financial Bid (Cover B): Opening – 20.03.2019 at 11.00 AM in the address mentioned above. (Venue is mentioned at the address mentioned above) (Bidders / authorized representative may remain present at the time of opening of bid) |

SECTION -II

IMPORTANT INSTRUCTIONS TO BE NOTED CAREFULLY BY THE TENDERERS

| | | |
|----|--|--|
| 1. | Mode of Procurement | Through Open Advertisement |
| 2. | Purchaser | Chief District Medical & Public Health Officer, Kandhamal |
| 3. | Consignee | DHH,SDH,AH,CHC etc of District Kandhamal |
| 4. | Delivery Period | Within 30days from issue of the purchase order. |
| 5. | Mode of Delivery | By Air / Road / Rail |
| 6. | Guarantee / Warranty /CMC | <u>Comprehensive warranty</u> including all spares, maintenance etc. for a period 3(<i>three</i>) years from the date of installation & commissioning and 3(<i>three</i>) years CMC after warranty period. |
| 7. | Tender Document Cost | Rs. 2,000/- : The tender document cost (Non-refundable) is to be submitted in the shape of bank draft in favour of the ZSS Non NRHM Fund, Kandhamal from any Nationalized / Scheduled Bank payable at Phulbani. |
| 8. | Earnest Money Deposit (EMD) (The approx. no. of equipment is mentioned in the Schedule of requirement – Section IV) | <p>The bidder may quote for any or all the equipment by submitting the required EMD (Refundable) of Rs 30,000/-.</p> <p>The Earnest Money Deposit will be paid in the shape of demand Draft only in favour of ZSS Non NRHM Fund, Kandhamal from any Nationalized / Scheduled Bank payable at Phulbani.</p> <p>EMD exemption is not permitted except to local SSI units registered in Odisha only as mentioned in Clause-4.3.</p> |

SECTION -III

TERMS & CONDITIONS

Sealed tenders are invited from the eligible bidders as per the eligibility criteria for **supply & installation of Equipment, Instrument, Furniture and Other Items** for the District, Kandhamal.

1. The bidders have to submit their tenders in separate sealed covers (i.e. **Cover "A"-Technical Bid & Cover "B"- Price Bid**). Both the covers should be put into a third **Cover "C"** which must be super-scribed as **"Tender for "Supply & installation of Equipment, Instrument, Furniture and Other Items"** and **Tender Reference No.** 3129 /DWH.

2. **The Cover "A" (Technical Bid) should contain as follows:**

- 1) Checklist with details of the documents enclosed in **Cover "A"** (as per **Annexure - I**) with page number. The document should be **serially arranged** as per this **Annexure - I** and should be securely tied and bound.
- 2) Manufacturing unit / supplier, who has been blacklisted either by the Tender inviting authority or by any state Govt. or Central Govt. organization, is not eligible to participate in the tender for that item during the period of blacklisting. They should submit a declaration about this by a **Notary Public**.
- 3) EMD **Rs.30,000/-** (Rupees Thirty Thousand only - Refundable) in shape of Bank Draft / Bankers Cheque in favour of ZSS Non NRHM Fund, Kandhamal payable at Phulbani.
- 4) Tender Paper Cost **Rs.2000/-** (Rupees Two Thousand only – Non-refundable) in shape of Bank Draft / Bankers Cheque in favour of ZSS Non NRHM Fund, Kandhamal payable at Phulbani.
- 5) List of Item (s) Quoted individually in the prescribed format (**Annexure – II**).
- 6) Copy of organization PAN.
- 7) Copy of organization Income Tax Acknowledgement Report (**Assessment Year 2016-17, 2017-18, 2018-19**) and copy of the audited financial statement for the last three financial year i.e. 2015-16, 2016-17 & 2017-18.
- 8) Photocopy of the GST registration certificate and last three months GST return filling copy i.e. Oct-2018 to Dec-2018.
- 9) The manufacturer / supplier should have 5 years market standing in supplying EIF to Govt./Corporate/PSU Hospitals in India. The copy of purchase order from the user should be furnished in support of the information provided in the market standing statement (item wise).

- 10) Performance Statement during the last three years towards proof of supply of similar EIF to any Govt./Corporate/PSU Hospitals in India. The copy of certificate from the user should be furnished in support of the information provided in the performance statement (item wise).
- 11) They should submit their valid drug license to the effect that they are dealing with the drugs, if they quoted for drugs.
- 12) Original Copy of Valid Manufacturing License of the manufacturer (s) / Import License by the Importer from the Original Equipment Manufacturer (OEM).
- 13) Copy of valid ISO Certificate.
- 14) Copy of valid ISI / CE / BIS / US FDA / IEC certificate.
- 15) All the tender documents should sign by the bidders at the bottom of each page with his official seal duly affixed.
- 16) Leaflet / Technical Brochures of the products / item offered.
- 17) They should quote the rates for individual items inclusive of packing, forwarding, door delivery and exclusive of GST and should submit a self declaration about this.
- 18) They must submit the undertaking that they will supply the stocks **within 30 days** after receipt of the Purchase Order from this office. In case of non-supply, the authority may allow extension for a **maximum period of 02 (two) weeks (14 days)** after the stipulated date of supply with a **penalty of 0.5% per week**, which will be deducted from the purchase order value as "**Liquidated Damage**".
- 19) If the supplier fails to complete the supply within the extended period, i.e. 44 days after being allowed by the purchaser, no further purchase order will be placed to the firm for the said item including forfeiture of the EMD and the concerned firm will be blacklisted for three (3) years from the date of issue of letter for the said item.
- 20) Violating the tender terms and conditions & non supply / supply which is not as per technical specification will disqualify the firm to participate in the tender for a period of three (3) years from the date of issue of letter and his E.M.D will be forfeited and no further purchase order will be placed to that firm for that item.
- 21) The supplier shall have a minimum Annual average turnover of Rs.1 (One) Crores or more in the last three year financial years i.e. 2015-16, 2016-17 & 2017-18 and copy of the audited financial statement for the last three financial year i.e. 2015-16, 2016-17 & 2017-18.

3. General Condition:

- 1) Eligible bidders should submit their tender documents to the CDM&PHO Kandhamal through **Speed Post / Registered Post / Courier only on or before 18.03.2019 by 5 PM.**

- 2) Any tender documents received after the due date & time will be rejected and returned to the sender unopened.
- 3) Violating the tender terms and conditions & non-supply / partially supply / supply, which is not as per technical specification, will be declared as **Blacklisted** and disqualify the firm to participate in any tender of this district for a period of next 03 (three) years from the date of issue of the letter and his E.M.D. will be forfeited.
- 4) Tenders documents should be typewritten or computerized, failing which the bidders will be ineligible for consideration. No further correction will be allowed.

4. The Cover "B" (Price Bid) should contain as follows:

- 1) Financial Bid must be submitted in the prescribed format as attached in **Annexure- IV**. No other document should be enclosed in the Financial Bid. The Financial Bid should be sent in a separate sealed cover called Cover "B" (Price Bid).
- 2) The rates should be computerized.
- 3) The price quoted by the tenderers shall not in any case, exceed the controlled price, if any, fixed by the Central / State Government / DGS&D and the Maximum Retail Price (MRP). The purchaser, at his discretion, will in such case, exercise the right of revising the price at any stage so as to confirm to the controlled price or MRP as the case may be deleted.
- 4) The rate quoted and accepted will be binding on the tenderer for a period of **one year** from the date of approval of the rate contract and on no account, any increase in the price will be entertained till the completion of this tender period.
- 5) If there is difference between figures & words, words will be taken into consideration.

The last date for submission of the tender documents is fixed on **18.03.2019 by 5 PM**. The Technical Bid will be opened on **19.03.2019 at 11 AM** by the purchase committee.

The undersigned reserves the right to reject or cancel any or all the tenders without assigning any reason thereof.

Chief District Medical & PH Officer
Kandhamal

CHECK LIST

| Sl. No. | List of Tender Documents Submitted | Yes | No | Page No. |
|---------|---|-----|----|----------|
| 1 | Cover "A"-Technical Bid | | | - |
| 2 | Cover "B"- Price Bid | | | - |
| 3 | EMD Rs.30,000/- (Rupees Thirty Thousand only - Refundable) in shape of Bank Draft / Bankers Cheque in favour of ZSS Non NRHM Fund, Kandhamal payable at Phulbani. | | | |
| 4 | Tender Paper Cost Rs.2000/- (Rupees Two Thousand only – Non-refundable) in shape of Bank Draft / Bankers Cheque in favour of ZSS Non NRHM Fund, Kandhamal payable at Phulbani. | | | |
| 5 | List of Item (s) Quoted in the prescribed format (Annexure – III) | | | |
| 6 | Self attested photocopy of organization PAN. | | | |
| 7 | Self attested photocopy of organization Income Tax Acknowledgement Report (Assessment Year 2016-17, 2017-18, 2018-19) and copy of the audited financial statement for the last three financial year i.e. 2015-16, 2016-17 & 2017-18. | | | |
| 8 | Self attested photocopy of the GST registration certificate and last three months GST return filling copy i.e. Oct-2018 to Dec-2018. | | | |
| 9 | Self attested photocopy regarding 5 years market standing in supplying above items to Govt./Corporate/PSU Hospitals in India. | | | |
| 10 | Performance Statement during the last three years towards proof of supply of similar EIF to any Govt./Corporate/PSU Hospitals in India. The copy of certificate from the user should be furnished in support of the information provided in the performance statement (item wise). | | | |
| 11 | They should submit their valid drug license to the effect that they are dealing with the drugs, if they quoted for drugs. | | | |
| 12 | Self attested photocopy of valid ISO certificate. | | | |
| 13 | Self attested photocopy of valid ISI / CE / BIS / US FDA / IEC certificate. | | | |
| 14 | All the tender documents should sign by the bidders at the bottom of each page with his official seal duly affixed. | | | |
| 15 | Leaflet / Technical Brochures of the products / item offered. | | | |
| 16 | Declaration by Notary Public regarding blacklisted of Manufacturer / supplier either by the Tender inviting authority or by any state Govt. or Central Govt. organization as per Annexure-V . | | | |
| 17 | Original copy of Valid Manufacturing License of the manufacturer (s) / Import License by the Importer from the Original Equipment Manufacturer (OEM) as per Annexure-VI . | | | |
| 18 | Self Declaration regarding quote the rates for individual items inclusive of packing, forwarding, door delivery and exclusive of GST charges as per Annexure-VII . | | | |
| 19 | Self Declaration regarding the rate quoted and accepted will be binding on the tenderer for a period of one year from the date of approval of the rate contract and on no account, any increase in the price will be entertained till the completion of this tender period as per Annexure-VIII . | | | |
| 20 | Undertaking regarding they will supply the stocks within thirty days after receipt of the Purchase Order from this office as per Annexure-IX . | | | |
| 21 | Declaration regarding the supplier shall have a minimum annual average turnover of Rs.1 (One) Crores or more in the last three year financial years i.e. 2015-16, 2016-17 & 2017-18 as per Annexure-X . The bidders shall submit the audited financial statement for the last financial year i.e. 2015-16, 2016-17 & 2017-18 for verification of turnover. | | | |

TENDER ITEMS

| Sl. No. | Name of the Item |
|---------|---|
| 1 | Alluminium Ladder |
| 2 | Cold Box & Ice Pack |
| 3 | Thermo Hygrometer |
| 4 | 3-way Stopcock |
| 5 | Allis Forceps |
| 6 | Ambu Bag(Paediatric Size) with Baby Mask |
| 7 | Artery Forceps, Straight, 160mm Stainless Steel |
| 8 | Basin Stand |
| 9 | Binocular Microscope |
| 10 | Blood Sample Mixer Rotor |
| 11 | CBC Machine 5 part with reticulocyte count |
| 12 | Centrifuge Machine |
| 13 | Cheatle Holder |
| 14 | Cheatle's Forceps |
| 15 | Chromogenic Coliform Medium without Membrane Filter (Himedia) |
| 16 | Cidex Tray Set (27" x 7" x 45") (SS) |
| 17 | Colorimeter |
| 18 | Cord Cutting Scissors, Blunt, Curved on Flat, 160 mm ss |
| 19 | Dental Probe |
| 20 | Dipsticks for Urine test for Protein and Sugar |
| 21 | Disposable Nale Cathetor |
| 22 | Dressing Forceps Spring Type,160mm, Stainless steel |
| 23 | Ear Speculum - Metallic, Dull finish |
| 24 | EDTA Crystal (Himedia) |
| 25 | Elbow Gloves |
| 26 | Elisa Dengue NS1 Ag Kit (Meril) |
| 27 | ESR Stand with Tube |
| 28 | Examination Lamp |
| 29 | Fire extinguisher |
| 30 | Flashlight/Tooth Box - Type - Pre - focused (4 cell) |
| 31 | Flexi Mask Adult 2mtr. |
| 32 | Flexi Mask Child 2mtr. |
| 33 | Foetoscope |
| 34 | Gauze Cutting Scissors Stright |
| 35 | Gluteraldehyde |
| 36 | Hb Electrophoresis System |
| 37 | HiCrome UTI Agar (Himedia) |
| 38 | Hydraulic Hand Pallet truck |
| 39 | Hypochlorite Solution (1ltr.) |
| 40 | Interdental Cleaning Aids |

| | |
|----|--|
| 41 | Iron Rack |
| 42 | Isopropyl Alcohol |
| 43 | Jobson - Horne - Probe |
| 44 | Kelly's Haemostat Forceps straight 140mm |
| 45 | Kidney Tray |
| 46 | Magnetic Stirrer With Hot Plate |
| 47 | Malaria Rapid Test Kit |
| 48 | Micro - Pipette |
| 49 | Microcentrifuge Tube (2ml) (Tarson) |
| 50 | Mouth Gag |
| 51 | Nasal Speculum (St. Claire's) * |
| 52 | Near Vision Chart |
| 53 | Nebulizer |
| 54 | Needle Holder (8") |
| 55 | Neuber Chamber |
| 56 | Otoscope |
| 57 | Oxygen Cylinder with Trolley |
| 58 | Oxygen Hood |
| 59 | Oxygen Mask (Adult) |
| 60 | Oxygen Mask (Paed) |
| 61 | Oxyset Neonate |
| 62 | Power drill machine for Orthopaedic |
| 63 | Rapid Test Kit for Hepatitis B & C |
| 64 | Robertson Cooked Meat Medium (Himedia) |
| 65 | Scalpel Holder with Blade |
| 66 | Sickling Test kit |
| 67 | Sims Retractor/Depressor |
| 68 | Slide Drying Rack |
| 69 | Snellen vision chart |
| 70 | Specimen Collection Bottle |
| 71 | Stitch Cutter |
| 72 | Stop Watch |
| 73 | TC- DC Count Apparatus |
| 74 | Tongue Depressor |
| 75 | Tris Buffer (Himedia) |
| 76 | Tunning Fork |
| 77 | Typhoid Rapid test kit |
| 78 | Universal Tooth Extractor |
| 79 | Uterine Sound Graduated |
| 80 | VIA Kit for Screening Cervical Cancer |
| 81 | Vulsellum Uterine Forceps Curved 25.5 cm |
| 82 | Weighing Scale for baby with 2 nos. of Sling |
| 83 | Yellow Tips for Micro - Pipette |
| 84 | Disposable Baby Dress Kit (two sets of Dress, Cap, Gloves, Socks, Napkin) (All Size) |

| | |
|--------------------------------|--|
| 85 | Disposable Surgeon Gown |
| 86 | Disposable Patient Gown |
| 87 | Disposable Draw Sheet |
| 88 | VIA Kit with Sterilizer |
| 89 | Equipments for VIA |
| 90 | OVE Kit |
| 91 | Equipments for OVE |
| 92 | Semi Recliner Bed (Fowers Beds for Mother) |
| 93 | Recline Chair (TechnicalKMC Chairs) |
| 94 | Bag & Mask (size 0 & 1) |
| MENTAL HEALTH MEDICINES | |
| 1 | Tab Heloperidol 5mg |
| 2 | Tab Epton 100mg |
| 3 | Tab Olanzapine 5mg |
| 4 | The Sodium Valporate 500mg |
| 5 | Tab Carbamazipine 300mg |
| 6 | Tab Amitriptylline 25mg |
| 7 | Tab Promethazine 25mg |
| 8 | Tab Trihexyphenidyl 2mg |
| 9 | Tab Nitrezepam 10mg |
| 10 | Tab Donepezil 5mg |
| 11 | Tab Fluxotine 20mg |
| ORTHPEDIC ITEMS | |
| 1 | Orthopedic power drill with K-wire (different sizes), reamer, drill bit attachment and oscilating saw attachment |
| 2 | Both bone forearm plating set |
| 3 | Humerous shaft plating set |
| 4 | Femur Nailing Set |
| 5 | Tibia Nailing Set |
| 6 | Distal Femur Locked Plating Operating set |
| 7 | Dynamic Hip Screw Operating Set |
| 8 | Proximal Femur Nailing Operating Set |
| 9 | Distal Radius Locking Plate Operating Set |
| 10 | Proximal Tibia plating operating set |
| 11 | K-wires of Different sizes |
| 12 | 3.5 DCP of different holes. |
| 13 | 4.5 DCP broad & narrow |
| 14 | Femur Antegrade interlocking nails of different sizes |
| 15 | Tibia Antegrade interlocking nails of different sizes |
| 16 | Dynamic Hip Screw both long & short barrel of different sizes |
| 17 | Proximal Femur Nails both long & short of different sizes |
| 18 | Distal Femur Locked Plating of different hole sizes |
| 19 | Proximal Tibia locking plates of different holes |
| 20 | Proximal Humorous locked plating operating set |

| | |
|----|---|
| 21 | Proximal Humerous locked plates of different sizes |
| 22 | Drill bits of different sizes |
| 23 | AO external fixators |
| 24 | 1/3rd Tubular plates |
| 25 | Selfretaining forceps of different sizes |
| 26 | Tissue cutting scissors (both fine & blunt) |
| 27 | Artery Forcep (Straight & Curved) of different sizes |
| 28 | Mosquito Artery Forcep |
| 29 | Allieys Forcep |
| 30 | Lead Apron |
| 31 | Lead Neck Guard |
| 32 | Gonad Guard |
| 33 | C-Arm Goggles |
| 34 | Steinmann Pins of different sizes |
| 35 | Schanz Pins of different sizes |
| 36 | JESS Fixator |
| 37 | Distal Radius Locking Plate |
| 38 | Letter Badged LR and O-9 |
| 39 | Pneumatic Tourniquet Set |
| 40 | Eshmark Tourniquet Upper & Lower limbs |
| 41 | Rush Nails of different sizes |
| 42 | Monopolar & bipolar cautery |
| 43 | Vacuum Closed wound suction sets of upper & lower limbs |
| 44 | Sterile Polytrolley Drape |
| 45 | Suction Cannula for upper & lower limbs |
| 46 | TUR Set |
| 47 | K-wire Cutter |
| 48 | K-wire Bender |
| 49 | Langenbach Retractor for upper & lower limbs |
| 50 | Homans Retractor for upper & lower limbs |
| 51 | Periosteum, Chisel, Osteotome for upper & lower limbs |
| 52 | SS-wire of 18, 22 sizes |
| 53 | Patellar Reduction Clamps |
| 54 | Pointed Reduction Clamps |
| 55 | Bulb Point |
| 56 | Towel clips (pointed & blunt) |
| 57 | Thumb forcep of different sizes (toothed & non-toothed) |
| 58 | Plyer |
| 59 | Depth Gauge for upper & lower limbs |
| 60 | Drill bits Sleeves for neutral & ascerntics |
| 61 | Bone tap of different sizes |
| 62 | Bone holding forcep for forearm, humerous, femur tibia |
| 63 | Steinmann Pins cutter |
| 64 | Different cancellous & cortical both locking & nonlocking for different plate sizes |

DENTAL EQUIPMENTS

| | |
|------|--------------------------------------|
| 1 | Dental Extraction Forcep Set |
| 1.1 | Anterior Forcep |
| 1.2 | Premolar Forcep |
| 1.3 | Molar Forcep |
| 1.4 | 3rd Moral Forcep |
| 1.5 | Cow horn Forcep |
| 1.6 | Bionate Forcep |
| 1.7 | Molar Root Forcep |
| 2 | Dental Extraction Elevator Set |
| 2.1 | Ohm's Periosteal |
| 2.2 | Coupland Elevator |
| 2.3 | Cryer Elevator |
| 3 | Pedontics Extraction Forcep |
| 4 | Mouth Mirror |
| 5 | Probe |
| 6 | Explorer |
| 7 | Dental Chair with Light System |
| 8 | Compressor for Dental Chair |
| 9 | Restorative Matterials |
| 9.1 | GIC |
| 9.2 | Composite Set |
| 9.3 | Zinc Oxide Eugenol |
| 9.4 | Amalgam |
| 10 | Scaler with tip for oral prophylaxis |
| 11 | Povidone Iodine Mouth Rinse |
| 12 | Chlorohexidine Solution 2% |
| 13 | Airoter with all burs |
| 14 | Endodontic Files |
| 15 | Root Canal Sealer |
| 16 | Gutthapercha |
| 17 | RC Irrigant (NaOCl) |
| 18 | Plastic Filling Instrument |
| 18.1 | Condenser |
| 18.2 | Ball Burnisher |
| 18.3 | Amalgam Carrier |
| 19 | Suction system with Suction tip |
| 20 | Prosthodontic Instruments |
| 20.1 | Bowl |
| 20.2 | Spatula |
| 20.3 | Impression Matterial (Alginate) |
| 20.4 | Dental Stone |
| 20.5 | Plaster of Paris |
| 20.6 | Heat & Cold cure resins |

| | |
|------|-----------------------------|
| 20.7 | Impression Trays |
| 20.8 | Prostodontic Lab Equipments |

LIST OF ITEMS QUOTED

| Sl. No. | Tender Quoted Sl. No. | Name of the Item (As per Annexure-II) | Manufacture Name | Make | Model Name | Specification submitted by the Bidders (mention details) |
|---------|-----------------------|---------------------------------------|------------------|------|------------|--|
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

Signature of the Bidder with seal

PRICE BID

| Sl. No. | Name of the Item | Make & Model | Unit Price with all accessories (as per Tender Clause No.-2.17) (both in words and figures) |
|---------|------------------|--------------|---|
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Signature of the Bidder with seal

DECLARATION

(Filled by the Notary)

I / We _____ do hereby declare that I / We have **not been de-recognized / black listed** by the Tender inviting authority or by any state Govt. or Central Govt. organization for supply of **Not of Standard Quality Items / non-supply**.

I / We agreed that the Tender Inviting Authority can forfeit the Earnest Money Deposit and blacklist me / us for a period of 3 years if, any information furnished by me / us proved to be false at the time of inspection / verification and not complying with the Tender terms & conditions.

Seal & Signature of the Notary

MANUFACTURER'S AUTHORISATION FORM

*(to be submitted by authorized distributor/importers in a **letterhead** in case the bidder is the authorized distributor/importer of OEM)*

No.

Dated:

To

**The Chief District Medical & Public Health Officer,
Kandhamal, Phulbani, Odisha**

Dear Sir / Madam,

Bid Reference No. :
Equipment Name :

We (name of the OEM) are the original manufacturers of the above equipment having registered office at (full address with telephone number/fax number & email ID and website), having factories at _____ and _____, do hereby authorize M/s. _____ (Name and address of bidder) as _____ (Importer / Distributor) to submit bids, and subsequently negotiate and sign the contract with you against the above bid no..

No company or firm or individual other than M/s. _____ are authorized to bid, negotiate and conclude the contract in regard to this business against this specific bid reference no.

We also hereby undertake to provide full guarantee/warranty /CMC/AMC as agreed by the bidder in the event the bidder is changed as the dealers or the bidder fails to provide satisfactory after sales and service during such period of Comprehensive warranty/CMC/AMC and to supply all the spares/reagents / consumables for 6 years.

We also hereby declare that we have the capacity to manufacture and supply, install and commission the quantity of the equipments bided within the stipulated time.

(Name)

for and on behalf of M/s. _____

(Name of manufacturers)

Date:

Place:

Seal

Note: *This letter of authority should be on the letterhead of the manufacturing concern and should be signed by a person competent and having the power of attorney to bind the manufacturer.*

DECLARATION

(Filled by the Bidder)

I / We _____ do hereby declare that I / We have quoted the rates for individual item inclusive of packing, forwarding & door delivery and exclusive of GST.

Signature of the Bidder with seal

DECLARATION

(Filled by the Bidder)

I / We _____ do hereby declare that the rate quoted and accepted will be binding on the tenderer for a period of **one year** from the date of approval of the rate contract and on no account, any increase in the price will be entertained till the completion of this tender period, otherwise the Tender Inviting Authority can **forfeit the Earnest Money Deposit and blacklist me / us for a period of next 3 years.**

Signature of the Bidder with seal

DECLARATION

(Filled by the Bidder)

I / We _____ do hereby declare that I / We supply the stocks **within 30 days** after receipt of the Purchase Order from the Tender Inviting Authority. In case of non-supply within the stipulated time period, the tender inviting authority may please be allowed extension for a **maximum period of 02 (two) weeks (14 days)** after the stipulated date of supply with a **penalty of 0.5% per week**, which will be deducted from the purchase order value as "**Liquidated Damage**".

I / We agreed that the Tender Inviting Authority can **forfeit the Earnest Money Deposit and blacklist me / us for a period of next 3 years for non-supply / part supply** of the stocks within the time period.

Signature of the Bidder with seal

ANNUAL AVERAGE TURN OVER STATEMENT
(To be furnished in the letter head of the Chartered Accountant)

The Annual Turnover of M/s _____ for the last 3 financial years are given below and certified that the statement is true and correct.

| Sl. No. | Financial Year | Turnover in Crore (Rs.) |
|--|----------------|-------------------------|
| 1 | 2015-16 | |
| 2 | 2016-17 | |
| 3 | 2017-18 | |
| Average Annual Turnover in Crore (Rs.) | | |

Date:

Signature of Chartered Accountant

Place:

(Name in Capital)

Seal Membership No.-

Note:

- 1) To be issued in the letter head of the Chartered Accountant with membership No.
- 2) Also attach photocopies of the audited P/L account of each year highlighting the turnover in support of that.

TECHNICAL SPECIFICATION OF THE ITEMS

1. ALUMINIUM LADDER:-

Step Ladder (8 or 10 Step Heavy Weight)

- Opening Size: 186.69 x 35.56 x 119.38 cm/Closing Size: 35.56 x 199.39 cm/Platform.
- Dimension: 25.4 x 25.4 cm/ Platform Height: 148.59 cm/Step Jump: 22.86 cm.
- Capacity is 150 kg / 330 lbs Heavy Duty Aluminium ladder for extra durability with a weight bearing capacity up to 150 kg, High quality aluminum and MUST have for every Home or Commercial Space
- Non-skid ribbed steps knee guard for added support Edge protector/saree guard anti-skid PVC shoes
- Ideal for every house hold, office, warehouse, godown or shop. Corrosion resistant structure for longer life. Wide base design for firm, comfortable and safe climb.

Simple Aluminium Ladder (8 or 10 Step)

- Material of Frame: High quality aluminium
- Surface Finish: oxidation
- Thickness: 1.2mm
- Max Load: 150kg
- High quality aluminum and MUST have for every Home or Commercial Space.
- Non-skid ribbed steps knee guard for added support Edge protector/saree guard anti-skid PVC shoes
- Ideal for every house hold, office, warehouse, godown or shop. Corrosion resistant structure for longer life. Wide base design for firm, comfortable and safe climb.

2. COLD BOX & ICE PACK:-

1. Cold Box is essential for safe transportation of sensitive vaccines from the place of manufacturing to the place of field storage and final carriage to the place of immunization. Thus, CFC-free cold boxes ensure a pre-determined safe temperature range (+2 to + 8 C) for a particular period known as cold life of the product. The cold life period varies according to the product classifications made by WHO i.e for large equipment like cold boxes, the cold life is high whereas for small equipment like vaccine carriers the cold life requirement is less. The Small cold box, long range should be able to keep the vaccine safe while transporting to long distances.

2. Technical Specifications

2.1 Vaccine Storage Capacity 5- 8 litres(any capacity within this range is acceptable 2.2 Weight Fully Loaded should be up to 35 Kgs

2.3 Weight Empty (with empty ice pack)should be up to 20 Kg

2.4 External surface and internal lining material LLDPE-(Linear Low Density Polyethylene)

2.5 Insulation material CFC-free Polyurethane

2.6 Insulation thickness 52-120 mm

2.7 Each Cold Box shall contain adequate Ice Packs of the following specifications: Volume of Ice Pack 0.25 to 0.4 litres. External dimensions 165*95*22 mm+/-1mm

2.8 Each vaccine carrier/cold box shall contain ice packs of WHO Specifications PIS Ref: E5/IP.2. The specifications are as given below: Water Content: 0.35 to 0.4 Litres. External Dimensions: 165x95x22 mm +/- 1 mm. Robustness: The Ice Packs samples shall withstand a one metre drop on every face, edge and corner when in a frozen state (-10 deg C to -20 deg C). It will then successfully pass the leakage test after thawing. Leakage Test: No leaks when 40 Kg lateral force is applied. Features: Effective reinforcements to restrain walls against swelling. Removable cap for filling ;cap to have effective internal water seal to resist 40 kg lateral force with no leakage. Manufacturer's recommended filling line to be clearly indicated. Freezing Characteristics: Shall not display super-cooling characteristics when filled with demineralised water and frozen in flat, horizontal contact with evaporator at - 8 deg C. Maximum thickness of the icepack, when frozen solid and laid flat on an evaporator surface, shall not exceed the unfrozen thickness by more than 25%. The internal dimension of the unit should be sufficient to accommodate the largest tolerances of the standard ice packs.

2.9 Should meet Cold Life Requirement as per WHO at 42 deg C without opening Minimum 48 hrs. 2.10 Should be provided with two handles to enable it to carried by one person

3. System Configuration Accessories, spares and consumables

3.1 Small Cold Box- Large Range(including integrated lockable fittings)- qty 3.2 Adequate Ice Packs to fill the box

4. Power Supply – None

3. THERMO HYGROMETER:-

- Fast response time
- High accuracy
- Simultaneous indication of temperature and humidity
- Automatic Stiwch-off
- Calculation of dew point and wet-bulb temperature
- Background lights
- Data memory of max. value
- Compact and light

Technical specifications

| | |
|----------------------|--|
| Measurement ranges | 0 to 100% r.h. -30 to 100°C |
| Resolution | 0.01% r.h . 0.01°C |
| Accuracy | ±2.0% r.h. at 25°C ±0.5°C at 25°C |
| Display | dual LCD with 4.5 positions |
| Battery | 9V battery (PP3) |
| Response time % r.h. | about 10 sec. (90 % @ 25° C air temp.) |
| Dimensions | 225 x 45 x 34mm |
| Weight | 200g |

9. BINOCULAR MICROSCOPE:-

- Universal power supply 150 to 220V (50 Hz) with electrical safety certificate.
- Rotatable head with skin-friendly soft touch acid resistant materials should be coated on Binocular piece, Co-axial coarse and fine adjustment knobs with vernier graduations.
- Quadruple revolving nose plane stage square with rack and pinion bidirectional 6V minimum 10 to 20W halogen lamp illuminator (dual).
- Abbe condenser NA 1.25 (oil) with variable iris diaphragm and blue filter.
- Oil Immersion.
- Binocular observation tube with standard inclination and inter-pupillary adjustment dioptre adjustment.
- Anti-fungus achromat 4x, 10x, 40x, 100x objectives parfocal.
- LB eyepiece 10x, FN 18mm.
- With dust cover, mirror and spare bulb.
- Comprehensive Warranty: 5 years inclusive thereafter 5 years CMC service support.

10. BLOOD SAMPLE MIXER ROTOR:-

- To be used to mix blood samples for CBC test.
- Simple rotor with rotational bars for mixing of sample tubes in slow speed up to 60 rpm, slightly slanting rotor bars to facilitate rotational and up down 3 dimensional rotary mixing.
- Power: AC 220V 50 Hz.
Comprehensive Warranty: 5 years inclusive thereafter 5 years CMC service support.

11. CBC MACHINE 5-PART WITH RETICULOCYTE COUNT:-

- Hydrodynamic focused impedance method, non-cyanide SLS hemoglobin method.
- Fluorescent flow cytometry based differential and reticulocyte counting with immature reticulocyte fraction count and reticulocyte hemoglobin equivalent count, automatic immature platelet fraction count, cumulative pulse height detection for HCT count.
- Optics: Semiconductor laser assembly with photodiode and PMT.
- High throughput.
- Separate reagent line for differential counting.
- Minimum sample volume: 25 µL (WB) to 70 µL (PD) modes.
- Memory: Data storage of 100000 samples with 10000 patient records.
- External printer with PC Compatible.
- USB interface with serial and LAN port.
- Optional bar code reader and autoloader.
- Power: 230 VAC, 50 / 60 Hz
- Comprehensive Warranty: 5 years inclusive thereafter 5 years CMC service support.

29. FIRE EXTINGUISHER:-

| Sl.No | Name of the Item | Specification | Unit |
|-------|----------------------------|---|------------------------|
| 1 | ABC type fire extinguisher | 1. Best quality. 2. Should be ABC type. 3. Should be 5/6 KG capacity 4. Should be IS – 2190 - 2010 | Each fire extinguisher |

| | | | |
|---|----------------------------|---|------------------------|
| 2 | CO2 type fire extinguisher | 1. Best quality. 2. Should be CO2 type. 3. Should be 4.5 KG capacity 4. Should be IS – 2190 - 2010 | Each fire extinguisher |
|---|----------------------------|---|------------------------|

36. Hb ELECTROPHORESIS SYSTEM:-

- Digital Electrophoresis power pack with 0 to 500V with 100V increment range 0 to 100 mA current with coarse and fine adjustments External and internal power fuse, timer, start on/off witch with polarity reverse facility.
- Smooth jack for supply connectors to tank Horizontal hard plastic tank with transparent lid, platinum wire and banana connectors.
- The tank should have an elevated middle portion with two buffer reservoirs below the level of the middle bridge to support non-submergible gel plate placement.
- Comprehensive Warranty: 5 years inclusive thereafter 5 years CMC service support.

38. Hydraulic Hand Pallet truck:-

- Capacity in kgs : 2500
- Min Fork Height in mm : 85/75
- Max Fork Height in mm : 195/185
- Lifting Height in mm : 110
- Fork Length in mm : 1150/1220
- Single Fork Width in mm : 160
- Width Over All Forks in mm : 550/685
- Fork Wheels in mm : 80*70
- Steering Wheels in mm : 180*50
- Self Weight in kgs : 86

Features:

- Technocart Hand Pallet Truck is built with heavy-duty components
- Constructed of extra thick steel with reinforced fork Stability
- Maintenance free
- 210° steering arc
- Spring loaded handle has 3 position – raise, lower, neutral with fingertip control

41. IRON RACK:-

- 6 shelves
- Height - 6 ft.
- Length - 3 ft. x 15 inch
- Shelves made off - 1 mm C.R. sheet
- Racks made off - 2 mm spotted angel

46. MAGNETIC STIRRER WITH HOT PLATE:-

- Digital temperature setting, simultaneous digital display of target and actual temperatures via LCD display, display of actual temperature value in medium with 0, 10C resolution.
- Heat Output (W): ~1500.
- Stirring Capacity: ~1.5L or better.
- Plate Dimensions (mm): 260 x 260 appx.
- Plastic material: Teflon coated.
- Power input / output (W): ~2 KW.
- Speed Range: 100-1500 rpm or better.
- Temperature range (0C): 50-200 or better.
- Light weight and table top model with user friendly model.
- Comprehensive Warranty: 5 years inclusive thereafter 5 years CMC service support.

57. OXYGEN CYLINDER WITH TROLLEY:-

- One Oxygen Cylinder (empty) (Refillable seamless steel gas cylinders) with valve and Cap. (BIS)-10 liters IS 7285 Part I II IS 3224 for neckring and Bull nose valve. The shoulder of the metal cylinder should carry a label stating Oxygen or Symbol O2 should be painted on the shoulder of the cylinder (painted as specified under Gas Cylinder Rules- 2004).
- Frame of the cylinder trolley is made with hi-grade MS tubular steel.
- Framework of the cylinder trolley mounted on two 10cm wheels.
- Finish in epoxy powder coated.
- For small (40cft) Oxygen Cylinder.
- Mounted on 100 / 150 mm diameter castor for mobility.
- Pre-treated and Epoxy Powder Coating for Scratch proof Finish.

62. POWER DRILL MACHINE FOR ORTHOPEDIC:-

Specification:-

| |
|--|
| Driving unit:- Includes motor, sturdy stand with wheels |
| Flexible shaft:- Minimum length-2 meters Autoclavable quick connection |
| Hand Piece for Drill:- Cannulated Autoclavable Pistol Type - Speed-1200 to 1500 RPM. - Jacob chuck. - Quick coupling chuck (Synthesis type). - Hudsons chuck. - Chuck for K-wire. - Forward & Reverse options. |
| Hand Piece for Reamers :- Cannulated Autoclavable Pistol Type - Speed-400 RPM, non damaging to the bone endosteal blood supply. - Chuck for cannulated Reamers. - Forward & Reverse options |
| Autoclavable Pistol Type - Easy Attachments of blades (without Instrument) - 2 Blades each of different size used routinely in Orthopaedic surgery (Total No, 10). - ACL Blades for commonly used sizes |

Environmental Factor:-

The unit shall be capable of being stored continuously in ambient temperature of 0 -50 deg C and relative humidity of 15-90%

The unit shall be capable of operating in ambient temperature of 20-30 deg C and relative humidity of less than 70%

Power Supply:-

Power input to be 220-240VAC, 50Hz fitted with Indian plug

Suitable UPS with maintenance free batteries for minimum one-hour back-up should be supplied with the system.

Resettable overcurrent breaker shall be fitted for protection

Standards, Safety & Training:-

Manufactures/Supplier should have ISO certificate to Quality Standard.

Should be compliant with IEC 61010-1:(or any international equivalent eg EN/UL 61010) covering safety requirements for electrical equipment for measurement control and laboratory use

Should be FDA, CE, UL or BIS approved product

Comprehensive training for lab staff and support services till familiarity with the system.

Proximal zigs should be of high quality , radiotransparent where needed and accurate in locking in safe zones of limb. Distal zigs if of good quality and accuracy will also be entertained.

Comprehensive warranty for 2 years and 5 years AMC after warranty

84. DISPOSABLE BABY DRESS KIT:-

- Each disposable baby dress kit contains two sets of dresses, caps, gloves, socks & napkins.

85. DISPOSABLE SURGEON GOWN:-

- 40 Gsm PPSB.
- Disposable.

86. DISPOSABLE PATIENT GOWN:-

- 25 Gsm PPSB.
- Disposable.

87. DISPOSABLE DRAW SHEET:-

- 30% Rayon / 70% Polyester
- Disposable
- Durable
- Comfortable on a patient's skin For small (40cft) Oxygen Cylinder.

88. VIA KIT WITH STERILIZER:-

This kit contains following items with specifications:

- Operation gloves (6.5", 7", 7.5")= 1 pkt of 100 gloves
- Cotton Swabs = 1 pkt of 100 swabs
- Distilled Water = 1 ltr. / bot
- Acetic Acid (1.5%) = 100ml 1 bot

89. EQUIPMENTS FOR VIA:-

This kit contains following items with specifications:

- Examination Lamp
- Cusco's Vaginal Speculum stainless steel 410 grade.
- Autoclave (small)
- Torch- Focusing torch with chrome / steel ribbed body (medium size) with battery.

90. OVE KIT:-

This kit contains following items with specifications:

- Wooden spatula / sticks
- Operation gloves (6.5", 7", 7.5")= 1 pkt of 100 gloves
- Cotton Swabs = 1 pkt of 100 swabs
- Gauze = Cotton gauze

91. EQUIPMENTS FOR VIA:-

This kit contains following items with specifications:

- Mouth Mirror (Angular mirror 2 pieces to provide indirect vision to retract lips, cheeks and tongue to reflect light into the mouth. Accurate image from flat surface mirror image magnifying with concave mirror size of 14mm – 22mm of plane size stainless steel, fog free with flat / round / ribbed handle. The mirror shall be made of fiber glass and the handle shall be made of stainless steel 410 / 420 grade.
- LED torch – LED focusing torch with chrome / steel ribbed body (medium size) with battery.

92. SEMI RECLINER BED (FOWLERS BEDS FOR MOTHERS):-

- Manufacturer should be ISO 9001 certified.
- Should have ISO 14001 – for environment friendly features.
- Should have ISO 18002 for occupational health safety management
- Product must be CE/BIS/BIFMA certified.
- Should furnish stainless steel grade certificate from Govt./ Govt. approved testing laboratory.
- Manufacturer should produce test certificate from Govt./ Govt. approval laboratory producer like impact test bend test, salt spray chamber test epoxy powder coating & phosphate coating for quoted item.

- Semi fowler bed made of Rectangular M.S.Tube with two sections top.
- The top should be made of CRCA perforated sheet.
- Back rest section has to be maneuvered by screw handle at foot and up to an angle of 60 degree.
- S.S. laminated head and foot boards.
- Bed mounted on 100mm Dia caster.
- Approximate dimension 200Lx90W60H.
- Finish should be pretreated & epoxy power coated.

93. RECLINE CHAIR TECHNICAL (KMC CHAIR):-

- It should be a folding gravity recliner chair.
- Smooth & easy reclining mechanism provides almost sleeping body posture.
- Unique & secured locking mechanism locks reclining position at any stage to provide desired body posture.
- Strong and stable design gives long lasting service.
- Easy to fold and unfold.
- Unique, strong capsule tune cross enables to take load upto 130kgs.
- Frame: powder coated round tubular mild steel.
- Gross weight: 8Kgs (Approximate).
- Dimension (approximate) seat 160 x 50Cms seat height 40 Cms.
- Manufacture should be ISO certified.

94. BAG & MASK (SIZE 0 & 1):-

Clinical Purpose:-

- To provide or assist ventilation in a patient who is apnoeic or exhibits inadequate respiration through manual pulmonary-driven pressure cycle functions.
- Used by clinical department/ward-it is used in ambulances intensive care units (ICU) during internal patients transfer, accident and emergency (A&E) and mass casualty incidents (MCI).

Overview of functional requirements

Technical specification:-

- Hand held manual resuscitator with transparent face mask.
- Child models (750ml, 500ml and 260ml bag capacity).
- Standard 15/22 mm swivel connector allows connection to all common masks Endo tracheal tubes.
- Provision to give supplementary oxygen-by-oxygen reservoir providing 100% oxygen.
- Non-re breathing valve enabling the patient to inspire oxygen from the reservoir bag.
- Should be single hand operatable.
- Should be easy to dissemble for cleaning and disinfection.
- Should have pressure release valve at 40cm H₂O.
- Should have silicone oxygen tube 2m length.
- It should be up to 40 times autoclavable including bag and washers.
- The bag should be made of soft silicone materials.
- Self inflating resuscitator bag should be of medical grade silicone rubber.
- The reservoir should be a PVC bag of 600ml capacity for 260ml & 500ml bag capacity and 1000ml for 750ml bag capacity.

- Accessories: silicone below, non Re-breathing valve. 2 meter oxygen tube, Guedel Airway, Oxygen Reservoir bag and Neonatal mask of 2 size viz 0 & 1.

Environment and Department considerations

Atmosphere/ Ambiance (air condition, humidity, dust...)

- Operating condition:- Capable of operating continuously in ambient temperature of 0 to 50 deg. C and relative humidity of 15 to 90% in ideal circumstances.
- An ambient air velocity is less than 0.3 m/s.

Users care cleaning, Disinfectant & sterility issues

Complete unit to be easily washable and sterilizable using both alcohol chlorine agents.

Standard and safety

- Certificate (Pre-market, sanitary...) Performance and safety standards (specific) to the device typed).
- Local and/or international-ISO 13485: Manufacturer/ supplier should have ISO certificate for quality standard.
- Should be USFDA/ CE (notified) approved product or BIS certified.
- Should meet ISO 10651-4 standard requirement.

Recommendations for maintenance – N/A

Notes:

- Service support contact details (Hierarchy wise: including a toll free/ landing number- NA).
- Recommendations or warning-NA.